




REVISIONS				
ZONE	REV.	DESCRIPTION	DATE	APPROVED



MODEL	TLTS P/N	CAPACITY (lbs-in)
12328-1K	160277	1,000
12328-2K	160280	2,000
12328-5K	160286	5,000

SPECIFICATIONS	
OUTPUT AT RATED TORQUE	2 mV/V NOMINAL
CALIBRATION ACCURACY	0.25% R.T.
OVERLOAD RATING	150%
EXCITATION MAX VAC OR VDC	20
BRIDGE RESISTANCE	350 Ω
ZERO BALANCE	± 1% F.S.
TEMPERATURE EFFECT ON R.T.	0.005% OF READING/°F
TEMPERATURE EFFECT ON ZERO	0.0025% R.T./°F
COMPENSATED TEMPERATURE RANGE	0°-150°F
CONNECTOR	MS3110P-8-4P

UNLESS OTHERWISE SPECIFIED:  DIMENSIONS ARE IN INCHES TOLERANCES ON: FRACTIONAL: ± 1/64 ANGULAR: ± 2 Degrees TWO PLACE DECIMAL: ± .010 THREE PLACE DECIMAL: ± .005  INTERPRET GEOMETRIC TOLERANCING PER:  MATERIAL 17-4 PH CONDITION H900  FINISH 125/✓				<div><div><div></div></div><div>TELEDYNE LECROY TEST SERVICES</div><div>Everywhere You Look</div><div>513 Mill St., Marion, MA 02738</div><div>Tel.: (508) 748-0103 Fax: (508) 748-2029</div></div> <div>TITLE:<div>REACTION TORQUE SENSOR ASSEMBLY</div></div> <div><div>SIZE B</div><div>DWG. NO. 12328-ASSY-1</div><div>REV 0</div></div> <div>SCALE: 1:2WEIGHT:SHEET 1 OF 1</div>
	NAME	DATE		
	DRAWN	KWL	6-5-17	
	CHECKED			
	ENG APPR.			
	Q.A.			
	PROJECT NO.			
	-			